



Certificate of Compliance

Nemko-CCL, Inc.

Certificate: NA201410320

Date Issued: February 7, 2014

Project: 249231-2.1

Issued to: VELP Scientifica S.r.l
Via Stazione
16 – 20865 Usmate (MB)
Italy

The products listed below have been certified as being compliant with all applicable requirements of the specifications listed and are eligible to bear the following certification mark



Issued by:

Robert Keller, Senior Engineer/Safety Supervisor

Authorized by:

Thomas Jackson, General Manager

PRODUCTS

MEASUREMENT, CONTROL, OR LABORATORY EQUIPMENT – Certified to US and Canada Standards

Product: Heating Digester and Lift Accessory

Model: Code F30100210; Code F30100200; Code F30100190; Code F30100180; Code F30150210; Code F30150190; Code F30150180; Code F30110200; Code F30110190; Code F30180200; Code A00000176; Code A00000170; Code A00000167

Ratings: All Class I equipment; Code F30100210, Code F30100180: 2300 W, 230 V~, 50-60 Hz; Code F30100200: 1150 W, 230 V~, 50-60 Hz; Code F30100190: 1500 W, 230 V~, 50-60 Hz; Code F30150210, Code F30150180: 2300 W, 200 V~, 50-60 Hz; Code F30150190: 1500 W, 200 V~, 50-60 Hz; Code F30110200: 1150 W, 115 V~, 50-60 Hz; Code F30110190: 1500 W, 115 V~, 50-60 Hz; Code F30180200: 1150 W, 100 V~, 50-60 Hz; Code A00000176, Code A00000170, Code A00000167: 50 W, 12 Vdc from Heating Digester

The certification system, as described in ISO/IEC Guide 67 (Conformity Assessment – Fundamentals of Product Certification), most closely resembles System 3

Nemko-CCL, Inc. 1940 West Alexander Street Salt Lake City, Utah 84119-2039 Tel (801) 972-6146 Fax (801) 972-8432



APPLICABLE REQUIREMENTS

UL Std. No. 61010-1 2nd Edition - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use – Part 1: General Requirements

CAN/CSA-C22.2 No. 61010-1-04 Second Edition - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use – Part 1: General Requirements

CAN/CSA-C22.2 No 61010-2-010-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use – Part 2-010: Particular Requirements for Laboratory Equipment for the Heating of Materials

This certificate is issued on condition that the holder complies and will continue to comply with the requirements of the above mentioned specifications and pursuant to the terms and conditions specified in the Certification Agreement.

Supplement to Certificate of Compliance

Certificate: NA201410320

Project: 249231-2.1

Nemko-CCL grants a license to the applicant to apply the Certification Mark to the certified products and that the mark shall only be affixed at the following factory locations

Factory Information

Factory Name	Location
VELP Scientifica S.r.l	Via Stazione 16 – 20865 Usmate (MB) Italy

The products listed, including the latest revision described below, are eligible to be marked in accordance with the referenced Certificate.

Product Certification History

Project	Date	Description
249231-2.1	February 7, 2014	Original Certification: Model: Code F30100210; Code F30100200; Code F30100190; Code F30100180; Code F30150210; Code F30150190; Code F30150180; Code F30110200; Code F30110190; Code F30180200; Code A00000176; Code A00000170; Code A00000167 Ratings: All Class I equipment; Code F30100210, Code F30100180: 2300 W, 230 V~, 50-60 Hz; Code F30100200: 1150 W, 230 V~, 50-60 Hz; Code F30100190: 1500 W, 230 V~, 50-60 Hz; Code F30150210, Code F30150180: 2300 W, 200 V~, 50-60 Hz; Code F30150190: 1500 W, 200 V~, 50-60 Hz; Code F30110200: 1150 W, 115 V~, 50-60 Hz; Code F30110190: 1500 W, 115 V~, 50-60 Hz; Code F30180200: 1150 W, 100 V~, 50-60 Hz; Code A00000176, Code A00000170, Code A00000167: 50 W, 12 Vdc from Heating Digester

This Supplement forms an integral part of the Certificate of Compliance

The certification system, as described in ISO/IEC Guide 67 (Conformity Assessment – Fundamentals of Product Certification), most closely resembles System 3

Nemko-CCL, Inc. 1940 West Alexander Street Salt Lake City, Utah 84119-2039 Tel (801) 972-6146 Fax (801) 972-8432

